



Ref. Certif. No.

SG PSB-MD-00884

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Adaptors (AC/DC ADAPTOR)
Name and address of the applicant	<b>ShenZhen RiHuiDa Power Supply Co., Ltd.</b> 2, 3 & 4F, No. B Building Rihuida Technology Park No. 28 Hongyuan Road, Pingdong Community Pingdi Sub-District Longgang District 518117 ShenZhen PEOPLE'S REPUBLIC OF CHINA
Name and address of the manufacturer	ShenZhen RiHuiDa Power Supply Co., Ltd. 2, 3 & 4F, No. B Building, Rihuida Technology Park, No. 28 Hongyuan Road, Pingdong Community, Pingdi Sub-District, Longgang District, 518117 ShenZhen, PEOPLE'S REPUBLIC OF CHINA
Name and address of the factory	ShenZhen RiHuiDa Power Supply Co., Ltd. 2, 3 & 4F, No. B Building, Rihuida Technology Park, No. 28 Hongyuan Road, Pingdong Community, Pingdi Sub-District, Longgang District, 518117 ShenZhen, PEOPLE'S REPUBLIC OF CHINA
Ratings and principal characteristics	Rated Input: 100-240VAC, 50/60Hz, 1.5A Rated Output: See test report for details Protection Class: II Degree of Protection: IPX0
Model/type Ref.	RHDzzfxxxxyy, RHDzzWxxxxyy, RHDzzDxxxxyy (See test report for details of model description)
Additional information (if necessary)	See Test Report for National Differences and Group Differences.
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1:2005/AMD2:2020
as shown in the Test Report Ref. No. which forms part of this certificate	211-27230036-000

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CBS 086019 0084 Rev. 00

Date, 2024-05-17

( Yager Bi )

TÜV SÜD PSB Pte Ltd 15 International Business Park TÜV SÜD @ IBP Singapore 609937



PSB Singapore